

# Aggregate Crushing Value Apparatus

ZI 6010



## Aggregate Crushing Value Apparatus : ZI 6010

IS 2376 (PART-IV) & BS 812.

For Measuring of Resistance of Aggregate to Crushing.

### SPECIFICATION:

Consists of M.S. Cylindrical Container 150mm +/-0.5mm dia x 130mm to 140mm high with base plate 200 to 230mm/sqr x 6mm thick. A Plunger of 148mm +/-0.5mm dia x 100 to 115mm high. Supplied complete with Tamping Rod, 16mm dia x 600mm long, both ends rounded, 1 no. Metal Measure 115 +/- 0.5mm dia x 180 +/- 0.5mm high.

### NOTE:

On Special Request Aggregate Crushing Value Apparatus having 75mm dia or 300mm dia cylinder size can be supplied.

## **ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA  
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558  
E-mail : sales@zealinternational.com ,website : zealinternational.com