


# Aggregate Impact Tester with Blow Counter

ZI 6011



## Aggregate Impact Tester with Blow Counter : ZI 6011

IS 2386 (PART IV) 9377 & BS 812.

For Determining The aggregate Impact Value.

### SPECIFICATION:

The instrument consists of a circular base with Two vertical guides. The Hammer of weight  $13.75 \pm 0.25$ kg can be raised to fall freely down the vertical guides. The Height of fall can be adjusted through  $380 \pm 5$ mm. The Hammer is provided with a locking arrangement. The hammer falls freely to the base and is removable for employing. Supplied complete with metal measure 75mm dia x 50mm high (For specimen preparation) and Tamping Rod 230mm long x 10mm diameter. A Blow Counter to count the number of strokes is fitted on top of the equipment.

## **ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA  
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558  
E-mail : sales@zealinternational.com ,website : zealinternational.com