

# Tile Abrasion Testing Machine

ZI 6018



## Tile Abrasion Testing Machine : ZI 6018

IS 1237, IS 1706

### SPECIFICATION:

This is used for determination of resistance to wear for Cement & Concrete flooring tiles. Tiles specimen of size 7.06 cm x 7.06 cm is pressed face-wise under specific load on a grinding path and abrasive powder is evenly spread on the rotating grinding disc the second parallel side of the tile is subjected to wear for similar number of rotations. The wear of the tile is measured on a thickness gauge specifically made for the purpose. The machine consists of a disc rotating at a speed of 30 R.P.M. in a circular tray. A bracket is provided to hold the specimen. A counter balance lever loads the specimen. Load applied is 30 Kgf. A funnel is fitted to evenly spread abrasive powder on the grinding path. A Pre-set Counter automatically stops the machine after 22 revolutions. This counter is Re-adjustable. The machine works on 440 V.A.C. Three Phase electrical supply. On Request machine to operate on 230 V.A.C. Supply can also be supplied.

## **ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA  
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558  
E-mail : sales@zealinternational.com ,website : zealinternational.com