

Hardness Tester for Mastic Asphalt

ZI 5005



Hardness Tester for Mastic Asphalt : ZI 5005

IS 1195

SPECIFICATION :

For determining the hardness number of Mastic Asphalt for flooring. It consist of an internally insulated cabinet to the base of which is fitted a water bath having two taps. The bath is heated by an immersion heater and the temperature is controlled with the thermostat at $35^{\circ} \pm 0.5^{\circ} \text{C}$. A 6.5mm dia pin is loaded on the specimen by means of a lever to give a 31.7 Kg weight. A dial gauge 0.01 x 25 mm is provided to record the penetration of pin into the specimen.

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558
E-mail : sales@zealinternational.com ,website : zealinternational.com