

Blaine's Air Permeability Apparatus

ZI 1002



Blaine's Air Permeability Apparatus : ZI 1002

IS 4031, 5516, 1727 & 4828, ASTM C-204, BS 4359-2

The apparatus is used for determining the fineness of cement in terms of specific surface expressed as total surface area in square centimeters per gram of cement. This is a variable flow type air permeability.

SPECIFICATION :

The apparatus consists one each of permeability cell 12.5mm I.D. manometer 'U' type mounted on stand with a built in stop cock, perforated disc, plunger rubber stopper, rubber tube 30cm long. Packet of 12 filter paper disc and a bottle of 100ml dibutylphthalate liquid.

SPARES & ACCESSORIES :

Punch to cut filter paper discs. Non-perforated disc. Suction bulb

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558
E-mail : sales@zealinternational.com , website : zealinternational.com