

Southard Viscometer

ZI 2004

Southard Viscometer : ZI 2004

IS 6932 (PART VI)

SPECIFICATION :

Putty consistency can be adjusted using this instrument. A 50mm I.D. metal cylinder is mounted on a wooden platform. A close fitting piston moves up to down without turning in this cylinder. Working stroke is 65mm. A metal bridge is provided to measure supplied complete with lime putty density vessel.

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558
E-mail : sales@zealinternational.com ,website : zealinternational.com