


Cone Penetrometer For Mortar

ZI 2008



Cone Penetrometer for Mortar : ZI 2008

IS 2250-1965

SPECIFICATION:

For determining the consistency of masonry mortar Consists of a movable bearing rod to which a cone 145mm. Long and 75mm dia at a base is fixed. The bearing rod passes freely through a bracket which is provided with release mechanism. A dial graduated in mm with rack and pinion is provided for measuring the penetration. Complete with a conical container 150mm id x 180mm deep and a platform.

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558
E-mail : sales@zealinternational.com ,website : zealinternational.com