

# Original Schmidt Test Hammer Type N

ZI 10001



## Original Schmidt Test Hammer Type N : ZI 10001

Measuring range 10 to 70 N/mm<sup>2</sup> compressive strength (below 25 N/mm<sup>2</sup> type P is better suited). Impact energy = 2,207 Nm. Rebound values are read from a scale. Testing the compressive strength of a prefabricated concrete girder. Rebound values are recorded by an assistant who will calculate mean values and read compressive strength values from a conversion diagram.

**ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA  
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558  
E-mail : sales@zealinternational.com ,website : zealinternational.com