

Pull-Off Tester DYNA

ZI 10010



Pull-Off Tester DYNA : ZI 10010

Testing is carried out directly on the structure. It can be measured on any point of the structure. For the determination of the tension strength it is not necessary to install test devices prior to the casting of the concrete. It is ideal for the evaluation of strength in concrete renovation.

ADVANTAGES

1. Mobile for use in any location
2. Independent of mains supply thanks to easy running crank drive
3. Constant jerk-free load increase
4. Weights only 3.5 kg (standard unit Z16)

MEASURING THE ADHESIVE STRENGTH OF APPLIED COATINGS

such as plastic coatings, concrete coats, mortars and plasters, bituminous coats and paint finishes and coatings on metal.

TECHNICAL INFORMATION

DYNA	Tensile force	Resolution	Accuracy	Permissible stroke
Z16	16 kN	0.01 N/mm ²	< 2%	3.5 mm
Z6	6 kN	0.01 N/mm ²	< 2%	4.0 mm

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558
E-mail : sales@zealinternational.com ,website : zealinternational.com