



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi - 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Length Comparator ZI 1014

Standards: IS 1199-1959, IS 4031 1968 BS 1881,  
ASTM C 151, C490.

### Specification:

The instrument consists of a channeled or a round base over which two vertical pillars are fixed. An adjustable cross plate is at the top. A dial gauge, reading to .002mm x 5mm. Can be located upon a 6.5mm. dia ball or other reference point cemented in the specimen. On the base there is similar recessed seating in which can be placed a second ball or reference point in the specimen. Complete with a stainless steel standardization bar with insulated grip and with 6.5mm dia. Balls mounted in the ends. The unit can be supplied with an Electronic Dial Gauge at extra cost if indicated at the time of placing the order.

