



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Devel Attrition Tester ZI 5021

Standards: ASTM D2-33 & IS 2386 (PART IV)

Specification:

It consists of two hollow cylinders closed at one end and provided with Fitting covers at the either end. These cylinders are mounted on a shaft at angle of 30° with the axis of rotation of the shaft. The shaft rotates at 30-33 RPM. Through a reduction gear operated by a motor and is provided with a revolution counter. Complete with Abrasive Charge consisting of 12 Nos. Hardened steel Balls of 48mm dia. Suitable for operation on 440 Volts, Three Phase, 50 Cycles, A.C. supply.  
Note : Option of Digital Preset Counter can be provided at an Extra Cost.

