

# Flash Point

ZI P-103



## **Flash Point : ZI P-103**

IS:1448 (Part 69), ASTM D92, IP 36 BS:4689, ISO 2592

For determining the flash and fire points of petroleum products, except fuel oils and those products having an open cup flash below 79°C (175°C)

### **ZI P 103**

Flash Point (Open) and Fire Point Cleveland, Electrical Heating with NPL Certificate

### **ZI P 104**

Flash Point (Open) and Fire Point Cleveland, Electrical Heating without NPL Certificate

### **The apparatus consists of:**

**ZI P 10301** Cleveland Flash Cup, brass, with an Insulated Handle

**ZI P 10302** Gas Test-Jet Assembly, correctly positioned and pivoted from a Cast Cup Platform which is fixed to a Heating Bath.

**ZI P 10303** Energy Regulator, to regulate the rate of rise in temperature, 220 V, 50Hz, Single Phase, AC Supply.

**Essential Extras** ZI P 10304 Thermometer IP 28 C, Range: -6° C to +110°C

## **ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA  
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558  
E-mail : sales@zealinternational.com ,website : zealinternational.com