


Abel Flash Point

ZI P-105



Abel Flash Point : ZI P-105

IS:1448 - Part 20, IS: 101, IP 170, IP 33, BS:3900

ZI P 105 Abel Flash Point Apparatus with NPL Certificate

Electrical Heating, with gas test jet and electric heater with ZI P 10502. Energy regulator. Suitable for operation on 220V, single phase, 50 Hz, AC supply. For determining closed cup flash point of petroleum products, their mixtures, other liquid and paints, having flash points between 19° C and 49° C. Each instrument bears a test certificate of the National Physical Laboratory & is supplied complete with an oil cup, cover fitted with gas test jet assembly, water bath and tripod stand, but without thermometer, in a wooden carrying case.

ZI P 106 : Abel Flash Point Apparatus same as ZI P 105 but without NPL Certificate.

Replacement Part with NPL Certificate

ZI P 10501 : Abel Oil Cup with lid, shutter and gas test jet brass.

Optional Extras

ZI P 10502 : Heating Element

ZI P 10503 : Abel oil cup thermometer, IP 74 Range: -35°C to + 70°C

ZI P 10504 : Abel water bath thermometer IP 75, Range: -30° C to + 80° C

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA

Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558

E-mail : sales@zealinternational.com ,website : zealinternational.com