

# Flash Point (Closed) Pensky-Martens Apparatus

ZI P-107



## **Flash Point (Closed) Pensky-Martens Apparatus : ZI P-107**

IS:1209, IS:1448(Part21), ASTM D 93, BS:2839, ISO 2719, DIN 51758, AFNOR M/07-019, Ip34

Flash Point Pensky Martens (Closed) apparatus with NPL Certificate, Electrical Heating, with gas test jet and electric heater with ZI P 10303. Energy regulator. Suitable for operation on 22V, single phase, 50 Hz, AC supply. Complete in wooden carrying case.

**ZI P 108** :Flash Point (Closed) Pensky-Martens Apparatus same as ZI P 107 but without NPL Certificate.

**ZI P 109** :Flash Point (Open) & Fire Point Pensky-Martens Apparatus with NPL Certificate.

Electrical heating, with gas test jet and electric heater with ZI P 10303 energy regulator. Suitable for operation on 220, 50 Hz, Single Phase, AC Supply. Complete in a wooden carrying case.

### **Essential Extras**

ZI P 10701 Thermometer IP 15C Range : -5° C to +110° C

ZI P 10702 Thermometer IP 16C Range: 90° C to +370° C

### **Replacement Parts**

ZI P 10703 Pensky-Martens Oil Cup

ZI P 10704 Pensky-Martens Lid shutter with Gas Test Jet

## **ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA

Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558

E-mail : sales@zealinternational.com , website : zealinternational.com