


Permeability Apparatus (Falling Head Permeability)

ZI 3026



Permeability Test Apparatus (Falling Head Permeability) : ZI 3026

IS 2720 (PART XVII) 1966.

The apparatus is used for the laboratory determination of permeability of soil using a constant or a variable head. This test is recommended for soils with co-efficient permeability in the range of $0 \cdot 10^{-3}$ to 10^{-7} cm/sec. The maximum particle size of the soil which can be tested in the mould is 10mm.

SPECIFICATION:

The equipment comprises one each :

Gun metal mould, 100mm I.D. x 127.3mm high x 1000ml Volume.

Gun metal mould extension collar, 100mm dia x 60mm high for the above mould.

Gun metal drainage base plate with a recess for a porous stone.

ZEAL INTERNATIONAL

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558
E-mail : sales@zealinternational.com ,website : zealinternational.com