

# IOWA Bearing Value Apparatus

ZI 3085

## IOWA Bearing Value Apparatus : ZI 3085

For determining the supporting strength of soil and aggregate mixtures which pass the 2mm and do not contain more than 10% sand size material.

The salient features of this apparatus are :-

- 1) The moulds used in this test are very small hence very smaller samples are required which is a great advantage when a large number of testes have to be carried out with a limited quality of soil available. Also when large quantities of soil are available fresh samples can be used for each test.
- 2) Time required this test is very much less than C.B.R. test and thus more number of tests can be carried out.

## **ZI 3086 IOWABEARING VALUE APPARATUS (MOTORISED)**

### **SPECIFICATION:**

Same as above but supplied with a bench mounting load frame motorized to give a constant strain rate of 1.25mm per minute.

## **ZEAL INTERNATIONAL**

1, Netaji Subhash Marg, Darya Ganj, New Delhi-110 002 INDIA  
Ph.:+91-11-23276114, 23244474, 23277528, 23278846, Fax :+91-11-23273214, 43580558  
E-mail : sales@zealinternational.com ,website : zealinternational.com