

ZEAL INTERNATIONAL (An ISO 9001 Certified Company) Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558 Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800 Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi - 110015

info@zealinternational.com sales@zealinternational.com support@zealinternational.com quotation@zealinternational.com /isit us at: www.zealinternational.com

Cement Sampler ZI 1001

Standards: IS 7535 1986, ASTM C 183, AASHTO T 127

Specification:

This is a brass tube approximately 53cm long and 2.8cm I.D. with a wooden handle. Total length approximately 73cm. The tube has the sharp angular edge which conveniently pierces cement bags. An air hole of approximately 3mm dia is drilled on the tube near handle. Total sample collected at one time 300 is cm

approximately.