

## Ring Mould ZI 1005

Standards: IS 712 1956, ASTM C 187

Specification:

Consists of mould 100 mm x 5mm. Deep and a glass base plates. Used to determine soundness of hydrated lime

