



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Aggregate Crushing Value Apparatus ZI 5014

Standards: IS 2386 (PART-IV) & BS 812:111

Specification:

The Aggregate Crushing Value (ACV) test set provides a relative measure of the resistance of an aggregate to crushing under a gradually applied compressive load. Each set consists of steel cylinder, plunger, base plate, cylindrical measure and tamping rod. All parts of the apparatus are powder coated or galvanized steel, heat treated and ground before manufacturing to make it durable and reliable.

Specification:

Consists of M.S. Cylindrical Container 150mm \pm 0.5mm dia x 130mm to 140mm high with base plate 200 to 230mm/sqr x 6mm thick. A Plunger of 148mm \pm 0.5mm dia x 100 to 115mm high. Supplied complete with Tamping Rod, 16mm dia x 600mm long, both ends rounded, 1 no. Metal Measure 115 \pm 0.5mm dia x 180 \pm 0.5mm high.

Note:

On Special Request Aggregate Crushing Value Apparatus having 75mm dia or 300mm dia cylinder size can be supplied.

