Semi Automatic Cone Penetrometer ZI 3005

Standards:BS 1377:2

Specification:

sed to determine the moisture content at which clay soils pass from a plastic to a liquid state. Two Version are available. One the standard one & the other is semi-automatic with timer where the cone is allowed to free fall for a period of 5 seconds including one each of 50 gm. & 100 gm. weight, one penetration cone, preset counter & measuring cup.

