



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi - 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Permeability Apparatus (Falling Head Permeability) ZI 3030

Standards: IS 2720 (PART XVII) 1966

Specification:

The apparatus is used for the laboratory determination of permeability of soil using a constant or a variable head. This test is recommended for soils with co-efficient permeability in the range of  $10^{-3}$  to  $10^{-7}$  cm/sec. The maximum particle size of the soil which can be tested in the mould is 10mm

