



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Sand Density Cone Apparatus ZI 3035

Standards:

Specification:

Used to determine the in-situ density of fine grained compacted soil. The test consist in digging a hole into the ground and then collect, dry and weight the sampled soil. The hole is then filled with dry sand from the cone container. The apparatus consist of a double metal cone, one plastic 5 lts jar & a metal tray with centre hole. Size of the cone is 6½" (165.1mm)

**Note:**

On request 12" (305 mm) Sand Density Cone Apparatus is also supplied at an extra cost.

