



**ZEAL INTERNATIONAL** (An ISO 9001 Certified Company)

**Head Office:** 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

**Branch Office:** 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

**Warehouse:** 4/26 Kirti Nagar Industrial Area New Delhi – 110015

[info@zealinternational.com](mailto:info@zealinternational.com)

[sales@zealinternational.com](mailto:sales@zealinternational.com)

[support@zealinternational.com](mailto:support@zealinternational.com)

[quotation@zealinternational.com](mailto:quotation@zealinternational.com)

Visit us at: [www.zealinternational.com](http://www.zealinternational.com)

## Earth Resistivity Meter ZI 3044

Standards: ASTM G 57

Specification:

- Capable of measuring earth voltage (V A.C.)
- 2mA measuring current permits the testing of earth resistance without tripping earth leakage current circuit breakers in the circuit.
- Auto power off function. The timer operates auto power off function. The time operates the "PUSH BUTTON SWITCH" and "TIMER ON BUTTON" are pressed together. This will keep the test "NO" for about three to five minutes.
- Then auto power off.
- Battery operated
- Data hold function
- Small and light weight
- IEC 1010 CAT III 200V
- Calibration performed with supplied test leads



### Specifications:

Measurement Ranges

Earth Resistance : 0 - 20W / 0 - 200W/0-2000W

Earth Voltage: 0-200V AC (40-500Hz)