



ZEAL INTERNATIONAL (An ISO 9001 Certified Company)

Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA

Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558

Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800

Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi – 110015

info@zealinternational.com

sales@zealinternational.com

support@zealinternational.com

quotation@zealinternational.com

Visit us at: www.zealinternational.com

Body Infrared Meter JPI 2013

Standards:

Specification:

Model	AT-150F
Measurement Range	32.0°C ~43.0°C
Accuracy	±0.3°C
Resolution	0.1°C
Measuring Distance	5cm~15cm
Operating Temperature	10~40°C
Relative Humidity	Operating:10%~90%RH
	Storing:1~80%RH
Power Supply	9V Alkaline Batteries
Size	88x42x175mm
Weight	120g (Not included Batteries)

