Bulk Destiny Apparatus JPI 7051

Standards:

Specification:

Bulk Density Apparatus: To determine the density of the powder graduated cylinder holded in clamps moves through motor up and down direction at a frequency rate between 28/32 cycles per minute through a distance of 5-6 cms. which strikes on a cushion fitted on the platform. A timer is provided to set the duration.

