

ZEAL INTERNATIONAL (An ISO 9001 Certified Company) Head Office: 1, Netaji Subhash Marg Daryaganj, New Delhi - 110002 INDIA Phone: +91-11-23276114, 23244474, 23250281, 23278846 Fax: +91-11-43580558 Branch Office: 4/20 Kirti Nagar Industrial Area New Delhi - 110015 Phone: +91 11 45380800 Warehouse: 4/26 Kirti Nagar Industrial Area New Delhi - 110015

info@zealinternational.com sales@zealinternational.com support@zealinternational.com quotation@zealinternational.com Visit us at: www.zealinternational.com

Ultrasonic Pulse Velocity Tester High Technology ZI 2075

Standards:

Specification:

- Measuring range: 0 3000 μs accuracy +/-0,1 μs
- Selection of the ultrasonic pulse amplitude adjustable from 250 to 1000 V
- Measurement of the required time by the ultrasonic pulse to go through the tested material.
- Single or continuous acquisition mode with automatic or manual saving.
- Zero calibration with depuration of the time for the pulse to go through the probes.
- Calibration of a defined time value.
- Capacity of data acquisition, processing and filing of the test data up to 30.000 samples.
- Interface mini USB for PC connection.
- Two outlets for connection to the oscilloscope.
- Languages: English, French, German, Spanish, Italian.

